

Code No. 27773

Mouse CCL8 / MCP-2 Assay Kit - IBL**INTRODUCTION**

Chemokine (C-C motif) ligand8/Monocyte chemoattractant protein-2 (CCL8/MCP-2) consists of 97 amino acids and is constitutionally-classified as C-C subfamily. CCL8/MCP-2 is produced by monocytes, fibroblasts and epithelial cells, and it shows chemotactic activity to CD4+ T cells, CD8+ T cells, monocytes and NK cells. Though the biological activity of CCL8/MCP-2 in mice is still unknown, it is recently reported that CCL8/MCP-2 in plasma increases with incidence of GVHD (Graft-Versus-Host-Disease) which is a complication of bone-marrow transplantation and the study suggests that CCL8/MCP-2 is associated with immune response within a living organism of mouse (ref. 1)

PRINCIPLE

This kit is a solid phase sandwich ELISA using 2 kinds of high specific antibodies. Tetra Methyl Benzidine (TMB) is used as a coloring agent (Chromogen). The strength of coloring is proportional to the quantities of Mouse CCL8/MCP-2.

MEASUREMENT RANGE

78.13 ~ 5,000 pg/mL

INTENDED USE

This IBL's assay kit is capable for the quantitative determination Mouse CCL8/MCP-2 in serum, EDTA plasma and Heparin plasma.

KIT COMPONENT

- | | | | |
|---|-------------------------------|--|------------|
| 1 | Precoated plate | : Anti-Mouse CCL8/MCP-2 (81) Rabbi IgG Affinity Purify | 96Well x 1 |
| 2 | Labeled antibody Conc. | | |
| | | : (30X) HRP conjugated Anti- Mouse CCL8/MCP-2 (33) Rabbit IgG Fab' Affinity Purify | 0.4mL x 1 |
| 3 | Standard | : Recombinant Mouse CCL8/MCP-2 | 0.5mL x 2 |
| 4 | EIA buffer | : 1% BSA, 0.05% Tween20 in PBS | 30mL x 1 |
| 5 | Solution for Labeled antibody | : 1 % BSA, 0.05 % Tween20 in PBS | 12mL x 1 |
| 6 | Chromogen | : TMB solution | 15mL x 1 |
| 7 | Stop solution | : 1N H ₂ SO ₄ | 12mL x 1 |
| 8 | Wash buffer Conc. | : (40X) 0.05 % Tween20 in phosphate buffer | 50mL x 1 |

OPERATION MANUAL**1. Materials needed but not supplied**

- | | |
|---|--------------------------------------|
| • Plate reader (450nm) | • Micropipette and tip |
| • Graduated cylinder and beaker | • Deionized water |
| • Refrigerator (as 4°C) | • Graph paper (log/log) |
| • Paper towel | • Tube for dilution of Standard |
| • Incubator (37°C ± 1°C) | • Washing bottle for precoated plate |
| • Disposable test tube for "2, Labeled antibody Conc." and "6, Chromogen" | |


2. Preparation

- Preparation of wash buffer
"8, Wash buffer Conc." is a concentrated (40X) buffer. Adjust the temperature of "8, Washing buffer Conc." to room temperature and then, mix it gently and completely before use. Dilute 50 mL of "8, Wash buffer Conc." with 1,950 mL of deionized water and mix it. This is the wash buffer for use. This prepared wash buffer shall be stored in refrigerator and used within 2 weeks after dilution.
- Preparation of Labeled antibody
"2, Labeled antibody Conc." is a concentrated (30X). Dilute "2, Labeled antibody Conc." with "5, Solution for Labeled antibody" in 30 times according to required quantity into a disposable test tube. Use this resulting solution as Labeled antibody.
Example)
In case you use one strip (8 well), the required quantity of Labeled antibody is 800 µL. (Dilute 30 µL of "2, Labeled antibody Conc." with 870 µL of "5, Solution for Labeled antibody" and mix it. And use the resulting solution by 100 µL in each well.)
This operation should be done just before the application of Labeled antibody.
The remaining "2, Labeled antibody Conc." should be stored at 4°C in firmly sealed vial.
- Preparation of Standard
Put just 0.5 mL of deionized water into the vial of "3, Standard" and mix it gently and completely. This solution is 10,000 pg/mL Mouse CCL8/MCP-2 standard.
- Dilution of Standard
Prepare 8 tubes for dilution of "3, Standard". Put 230 µL each of "4, EIA buffer" into the tube.
Specify the following concentration of each tube."

Tube-1	5,000 pg/mL
Tube-2	2,500 pg/mL
Tube-3	1,250 pg/mL
Tube-4	625 pg/mL
Tube-5	312.5 pg/mL
Tube-6	156.25 pg/mL
Tube-7	78.13 pg/mL
Tube-8	0 pg/mL (Test Sample Blank)

Put 230 µL of Standard solution into tube-1 and mix it gently. Then, put 230 µL of tube-1 mixture into tube-2. Dilute two times standard solution in series to set up 7 points of diluted standard between 5,000 pg/mL and 78.13 pg/mL. Tube-8 is the test sample blank as 0 pg/mL.

See following picture.

**5) Dilution of test sample**

Test sample may be diluted with "4, EIA buffer" as necessary.

200 - 500 multiplication is recommended as a rough guide for dilution of serum or plasma samples in normal mice. However, the dilution rate should be optimized by each laboratory since the level of Mouse CCL8/MCP-2 may vary with strain or immunological status of individual.

3. Measurement procedure

All reagents shall be brought to room temperature approximately 30 minutes before use. Then mix it gently and completely before use. Make sure of no change in quality of the reagents. Standard curve shall be prepared simultaneously with the measurement of test samples.

	Test Sample	Standard	Test Sample Blank	Reagent Blank
Reagents	Test sample 100 µL	Diluted standard (Tube 1~7) 100 µL	EIA buffer (Tube-8) 100 µL	EIA buffer 100 µL
Incubation for 60 minutes at 37°C with plate lid				
Washing 7 times				
Labeled Antibody	100 µL	100 µL	100 µL	-
Incubation for 30 minutes at 4°C with plate lid				
Washing 9 times				
Chromogen	100 µL	100 µL	100 µL	100 µL
Incubation for 30 minutes at room temperature (shielded)				
Stop solution	100 µL	100 µL	100 µL	100 µL
Read the plate at 450nm against a Reagent Blank within 30 minutes after addition of Stop solution.				

- Determine wells for reagent blank. Put 100 µL each of "4, EIA buffer" into the wells.
- Determine wells for test sample blank, test sample and diluted standard. Then, put 100 µL each of test sample blank (tube-8), test sample and dilutions of standard (tube-1-7) into the appropriate wells.
- Incubate the precoated plate for 60 minutes at 37°C after covering it with plate lid.
- Wash each well of the precoated plate vigorously with wash buffer using the washing bottle. Then, fill each well with wash buffer and leave the precoated plate laid for 15-30 seconds. Remove wash buffer completely from the precoated plate by snapping. This procedure must be repeated more than 7 times. Then, remove the remaining liquid from all wells completely by snapping the precoated plate onto paper towel.
In case of using a plate washer, after 4 times washing with plate washer, washing with above washing bottle must be repeated 3 times.
- Pipette 100 µL of labeled antibody solution into the wells of test samples, diluted standard and test sample blank.
- Incubate the precoated plate for 30 minutes at 4°C after covering it with plate lid.
- Wash the precoated plate 9 times in the same manner as 4).
- Take the required quantity of "6, Chromogen" into a disposable test tube. Then, pipette 100 µL from the test tube into the wells. Please do not return the rest of the test tube to "6, Chromogen" bottle to avoid contamination.
- Incubate the precoated plate for 30 minutes at room temperature in the dark. The liquid will turn blue by addition of "6, Chromogen".
- Pipette 100 µL of "7, Stop solution" into the wells. Mix the liquid by tapping the side of precoated plate. The liquid will turn yellow by addition of "7, Stop solution".
- Remove any dirt or drop of water on the bottom of the precoated plate and confirm there is no bubble on the surface of the liquid. Then, run the plate reader and conduct measurement at 450 nm against a reagent blank. The measurement shall be done within 30 minutes after addition of "7, Stop solution".

SPECIAL ATTENTION

- Test samples should be measured soon after collection. For the storage of test samples, store them frozen and do not repeat freeze/thaw cycles. Thaw the test samples at a low temperature and mix them completely before measurement.
- Test samples should be diluted with "4, EIA buffer", if the need arises.
- Duplicate measurement of test samples and standard is recommended.
- Use test samples in neutral pH range. The contaminations of organic solvent may affect the measurement.
- Use only wash buffer contained in this kit for washing the precoated plate. Insufficient washing may lead to the failure in measurement.
- Remove the wash buffer completely by tapping the precoated plate on paper towel. Do not wipe wells with paper towel.
- "6, Chromogen" should be stored in the dark due to its sensitivity against light.

"6, Chromogen" should be avoided contact with metals.


- 8) Measurement should be done within 30 minutes after addition of "7, Stop solution".

CALCULATION OF TEST RESULT

Subtract the absorbance of test sample blank from all data, including standards and unknown samples before plotting. Plot the subtracted absorbance of the standards against the standard concentration on log-log graph paper. Draw the best smooth curve through these points to construct the standard curve. Read the concentration for unknown samples from the standard curve.

Example of standard curve

Conc. (pg/mL)	Absorbance (450nm)
5,000	2.337
2,500	1.165
1,250	0.561
625	0.289
312.5	0.143
156.25	0.072
78.13	0.041
0 (Test Sample Blank)	0.005



* The typical standard curve is shown above. This curve can not be used to derive test results. Please run a standard curve for each assay.

PERFORMANCE CHARACTERISTICS

1. Titer Assay (Samples with standard added are used.)

Specimen	Titer (X)	Measurement Value (pg/mL)	Theoretical Value (pg/mL)	%
Mouse Serum (BALB/c)	100	3,393	3,275	103.6
	200	1,510	1,502	100.5
	400	705	705	100.0
	800	317	319	99.4
Mouse Plasma (EDTA) (BALB/c)	100	2,577	2,401	107.3
	200	1,168	1,087	107.5
	400	526	518	101.5
	800	243	244	99.6

2. Added Recovery Assay

Specimen	Theoretical Value (pg/mL)	Measurement Value (pg/mL)	%
Mouse Serum (BALB/c) (x200)	1,900	2,086	109.8
	1,572	1,655	105.3
	1,416	1,420	100.3
	1,338	1,320	98.7
	1,299	1,217	93.7
Mouse Plasma (EDTA) (BALB/c) (x200)	1,499	1,559	104.0
	1,022	1,101	107.7
	865	910	105.2
	787	784	99.6
	748	735	98.3

3. Intra - Assay

Measurement Value (pg/mL)	SD value	CV value (%)	n
2,691	182	6.8	20
526	39	7.4	20
250	23	9.2	20

4. Inter - Assay

Measurement Value (pg/mL)	SD value	CV value (%)	n
2,553	217	8.5	12
505	44	8.7	12
240	22	9.2	13

5. Specificity

Compound	Cross Reactivity
Mouse CCL8/MCP-2	100 %
Human CCL8/MCP-2	≤0.1 %
Mouse CCL2/MCP-1	≤0.1 %
Mouse CCL7/MCP-3	≤0.1 %

6. Sensitivity

49.9 pg/mL

The sensitivity for this kit was determined using the guidelines under the National Committee for Clinical Laboratory Standards (NCCLS) Evaluation Protocols. (National Committee for Clinical Laboratory Standards Evaluation Protocols, SC1, (1989) Villanova, PA: NCCLS.)

PRECAUTION FOR INTENDED USE AND/OR HANDLING

- All reagents should be stored at 2 - 8°C. All reagents shall be brought to room temperature approximately 30 minutes before use.
- "3, Standard" is lyophilized products. Be careful to open this vial.
- "7, Stop solution" is a strong acid substance. Therefore, be careful not to have your skin and clothes contact "7, Stop solution" and pay attention to the disposal of "7, Stop solution".
- Dispose used materials after rinsing them with large quantity of water.
- Precipitation may occur in "2, Labeled antibody Conc.", however, there is no problem in the performance.
- Wash hands after handling reagents.
- Do not mix the reagents with the reagents from a different lot or kit.
- Do not use expired reagents.
- This kit is for research purpose only. Do not use for clinical diagnosis.

STORAGE AND THE TERM OF VALIDITY

Storage Condition : 2 - 8°C

The expiry date is specified on outer box.

REFERENCE

- Hori T, Naishiro Y, Sohma H, Suzuki N, Hatakeyama N, Yamamoto M, Sonoda T, Mizue Y, Imai K, Tsutsumi H, Kokai Y. CCL8 is a potential molecular candidate for the diagnosis of graft-versus-host disease. Blood. 2008 Apr. 15;111(8):4403-12.
- Ota A, Yamamoto M, Hori T, Miyai S, Naishiro Y, Sohma H, Maeda M, Kokai Y. Upregulation of plasma CCL8 in mouse model of graft-vs-host disease. Exp Hematol. 2009 Apr;37(4):525-31.

Version 2.4